



## SCS Directory

Accreditation number: SCS 0159

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

Aquametro Oil & Marine AG  
Validation & Calibration  
Service (VCS)  
Ringstrasse 75  
4106 Therwil

Head: Marion Heusser  
Responsible for MS: Marion Heusser  
Telephone: +41 61 725 4416  
E-Mail: [mhs@aquametro-oil-marine.com](mailto:mhs@aquametro-oil-marine.com)  
Internet: [www.aquametro-oil-marine.com](http://www.aquametro-oil-marine.com)  
Initial accreditation: 05.01.2021  
Current accreditation: 05.01.2026 bis 04.01.2031  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

### Scope of accreditation as of 05.01.2026

#### Calibration laboratory for flow

##### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
Flow volume	0,003 m <sup>3</sup> /h ... 30 m <sup>3</sup> /h	Cold water	0.2 %	Scale
Flow volume		Gravimetric		
		Test oil		Oil density: 814 kg/m <sup>3</sup> at 20 °C Viscosity: 5 cSt at 20 °C
	0,5 l/h ... 200 l/h	Volumetric	0,06 %	Piston
	0,015 m <sup>3</sup> /h ... 30 m <sup>3</sup> /h	Volumetric	0,12 %	Mastermeter

In case of contradictions in the language versions of the directories, the German version shall apply.

\* / \* / \* / \* / \*